

Sterilization Procedure *

Products: 404 – Mouth Rest
413 –Dappen Ring

Sterilization by autoclaving:

Submit the product by autoclaving at 121 ° C and 1 atm for 30 minutes.**

Sterilization oven:

Submit the product in an oven at 170 ° for 120 minutes.**



Gyselle Lima
Coordinator of the Quality Management System.

* Refer to the autoclave manufacturer's manual or oven to be used.

** Conditions of temperature and pressure are related to sea level used in Brazil.